



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/765,318	01/26/2004	Keith G. Lurie	016354-005212US	9412

20350 7590 08/25/2005

TOWNSEND AND TOWNSEND AND CREW, LLP
TWO EMBARCADERO CENTER
EIGHTH FLOOR
SAN FRANCISCO, CA 94111-3834

EXAMINER

DOSTER GREENE, DINNATIA JO

ART UNIT	PAPER NUMBER
----------	--------------

3743

DATE MAILED: 08/25/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/765,318

Applicant(s)

LURIE, KEITH G.

Examiner

Dinnatia Doster-Greene

Art Unit

3743

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 11 September 2003.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-22 is/are pending in the application.
- 4a) Of the above claim(s) 15-20 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-6, 8, 9, 11-13, 21 and 22 is/are rejected.
- 7) ☒ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☐ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☒ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____.
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____.
- ☐ Notice of Informal Patent Application (PTO-152)
- ☐ Other: _____.

DETAILED ACTION

Election/Restrictions

Restriction to one of the following inventions is required under 35 U.S.C. 121:

- I. Claims 1-14 and 21-22, drawn to method for modulating a person's autonomic function, classified in class 128, subclass 203.11.
- II. Claims 15-20, drawn to a device for modulating a person's autonomic function, classified in class 261, subclass Dig. 65.

The inventions are distinct, each from the other because of the following reasons:

Inventions Group I and Group II are related as process and apparatus for its practice. The inventions are distinct if it can be shown that either: (1) the process as claimed can be practiced by another materially different apparatus or by hand, or (2) the apparatus as claimed can be used to practice another and materially different process. (MPEP § 806.05(e)). In this case, the method can be used with a device which does not include a housing having an opening.

Because these inventions are distinct for the reasons given above and have acquired a separate status in the art as shown by their different classification, restriction for examination purposes as indicated is proper.

During a telephone conversation with Darin Gibby on August 19, 2005 a provisional election was made without traverse to prosecute the invention of Group I, claims 1-14 and 21-22. Affirmation of this election must be made by applicant in replying to this Office action. Claims 15-20 are withdrawn from further consideration by the examiner, 37 CFR 1.142(b), as being drawn to a non-elected invention.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 4-9, 11-14 and 21 are rejected under 35 U.S.C. 102(b) as being anticipated by Alferness (U.S. Patent No. 4,349,015). Alferness in a breathing device discloses:

- Interfacing a valve system to the person's airway (22,26, 42),
- The valve system being configured to decrease or prevent respiratory gas flow to the person's lungs during at least a portion of an inhalation event (Abstract),
- Permitting the person to inhale and exhale through the valve system (22, 26, 42), wherein during inhalation the valve system functions to produce a vacuum within the thorax (col. 5, lines 28-68 and col. 6, lines 1-56). With respect to the language of "to transiently decrease intrathoracic pressure and thereby modulate the person's autonomic function" it would be inherent that the system of Alferness performs the functions because the

Art Unit: 3743

intrathoracic pressure is decreased during the bellow decompression (col. 3, lines 18-22).

With respect to claims 4-7, 8, 9, 11, 12-14, Alferness discloses that it device relates to controlling a patient's pulmonary functions (Alferness, entire document).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

Claims 2, 3, 10 and 22 are rejected under 35 U.S.C. 103(a) as being unpatentable over Alferness. As discussed above, Alferness discloses the claimed invention with the exception the specific actuating pressures claimed by Applicant. The Office takes the position that it would have been obvious to one of ordinary skill in the art at the time of the invention to arrive at the claimed range, based on the teachings of Alferness (col. 6, lines 41-56) "that those skilled in the art will accordingly recognize that

Art Unit: 3743

the predetermined value of pressure that is required to open the valve may be empirally varied to fit a specific design and opening mode of the apparatus." Also based on the individual which the valve system is to be employed, the actuating pressure of the valve would be varied based on patient criteria such as age of the user, disease process, disease to be treated because these factors would increase/decrease the actuating pressure needed in the valve system to provide appropriate ventilation which would be obvious to one of ordinary skill in the art.

Double Patenting

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and, *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 1-3 and 21-22 are rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claim 42 of U.S. Patent No. 6,604,523. Although the conflicting claims are not identical, they are not patentably distinct from each other because **claim 1 of the instant application**, limitations can be found in claim 42 of patent '523. The patented claim 42 is more specific than the instant application. Therefore patented claim 42 "anticipates" claims 1-3 and 21-22.

Allowable Subject Matter

Claims 7 and 10 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Dinnatia Doster-Greene whose telephone number is 571-272-7143. The examiner can normally be reached on 8:30-4:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Henry Bennett can be reached on 571-272-4791. The fax phone number for the organization where this application or proceeding is assigned is 571-273-7143.


Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Application/Control Number: 10/765,318

Page 7

Art Unit: 3743

ddg



Henry Bennett
Supervisory Patent Examiner
Group 3700